

## SAFETY PRECAUTIONS

- ◆ The instrument apply to general industrial occasions, but keep protect the device separately when you has the special using purpose. Do not operate in explosive atmosphere but keep vantilate on the installation site.
- ◆ Use Rated voltage of the power supply, correct wiring, Proper grounding, do not touch the terminals on back of the device to keep of electric shock.
- ◆ In order to be safty, keep out of charged when installation, maintenace and remove.
- ◆ It need take shielding steps when the device is close to the occassions of Mains power, high electric field, high magnetic fields and Ac contactor .
- ◆ In order to avoid measuring error when the sensor is thermocouple pls use the compensating wire correspondingly, when the sensor is thermal resistance need use three sane copper conductor and the resistance is less than 10Ω.
- ◆ Use the soft cloth to clean the instrument but keep out of organic solvents such as alcohol and petro; Turn of the power, stop using when the instrument is in water.
- ◆ In order to keep machine well, regular maintenance and maintenance is needed
- ◆ When open the cartons the instrument is broken by transportaion pls contact with the manufaturer.
- ◆ When open the cartons pls check the packing list carefully as follows:

No.	Name	Unit	QTY	NOTE
1	Paperless recorder	pcs	1	
2	User Manual	book	1	
3	Mounting bracket (with screws)	pcs	2	
4	Application	set	1	
5	Standard Software (CD)	set	1	
6	U disk	pcs		optional
7	RS232C lines	pcs		optional
8	RS232C/485 convert modula	pcs		optional
9	RS232C/CAN convert modula	pcs		optional
10	Micro printer (accessory)	pcs		optional

## CONTENTS

CHAPTER 1 TECHNICAL INDEX .....	4
CHAPTER 2 INSTALLATION AND WIRING.....	8
2.1 INSTRUMENT DIMENSION.. .....	8
2.2 TERMINAL CONNECTION.....	8
2.3 AC/DC POWER INPUT TERMINAL.....	8
2.4 INPUT/OUTPUT SIGNAL CONNECTION.....	9
2.5 TRANSMITTER CONNECTION.....	9
2.6 RELAY CONTACT OUTPUT TERMINAL.....	10
2.7 COMMUNICATION INTERFACE DEFINITION AND CONNECTION.....	10
CHAPTER 3 USAGE PATTERS AND OPERATION .....	11
3.1 KEYBOARD FUNCTION.....	11
3.2 RUN MODE.....	11
3.3 OVERVIEW SCREEN.....	12
3.4 DIGITAL DISPLAY .....	12
3.5 TREND SCREEN.....	13
3.6 HISTORY SCREEN.....	13
3.7 PRINT SCREEN.....	14
3.8 ALARM INFORMATION.....	15
3.9 POWER INFORMATION.....	15
3.10 BACKUP HISTORY DATA.....	16
CHAPTER 4 CONFIGURATION AND ASISTING OPERATION.....	17
4.1 CONFIGURATION AND SYSTEM CONFIGURATION.....	17
4.2 COMMUNICATION CONFIGURATION .....	18
4.3 INPUT CONFIGURATION.....	19
4.4 OUTPUT CONFIGURAION.....	21
4.5 SECONDARY INTERFACE.....	22

## CHAPTER 1 TECHNICAL INDEX

### ■ SCREEN

**SCREEN:**128\*64 dot matrix blue LCD screen

**ACCURACY:** REAL TIME DISPLAY:±0.2% F.S.  
RECALL ACCURACY:±0.2% F.S.

**Note:**Thermocouple should minus the cold end error

### ■ PROCESSOR

Using high-performance ARM Cortex-M3 32 bit RISC cores  
can realize multi-channel collection, record, display and alarm  
at same time.

### ■ MEMORY MODULA

Use large parallel capacity NAND FLASH MEMORY chip to  
storage history data, using serial FRAM chip to storage key  
system configuration parameters.

### ■ INPUT FUNCTION

**INPUT SPECIFICATIONS:**Isolate universal input, support  
Max 4 channel analog input.

**INPUT VOLTAGE:**0-5V、1-5V、0-20mV、0-100mV

**CURRENT INPUT:**0-10mA、4-20mA (0-20mA Noted when  
place order)

**INPUT RESISTANCE:**Res

**INPUT FREQUENCY:**FREQUENCY SIGNAL (PI) (frequency  
range: 0.5-5000HZ, The other can be customsized)

**THERMAL RESISTANCE:**PT100、Cu50、G53、Cu100、BA1、  
BA2 ( require three balance wire resistance, LEAD  
RESISTANCE<10Ω)

**THERMOCOUPLE:**S、B、K、T、R、E、N、J

**RADIATION PYROMETER:**F1、F2

**RHENIUM:**WRe3-25、WRe5-26

**Note:**Other signal input (such as switch value input (DI))  
or index number (such as PT1000) need noted when place  
order.

### ■ Out put function

**POWER DISTRIBUTION:**Transmitter Concentration

distribution +24VDC, standard distribution <30mA (MAX load  
capacity can be customized) ,Supports a variety of  
specifications and concentration distribution (such as 12VDC、  
5VDC)

**ANALOG OUTPUT:** Max support 2 channel 4-20mA standar  
current analog output, load capacity 750Ω (max), convenient  
for collection of indicating instrument or DCS/PLC, realize the  
long signal transmitter

**OUTPUT OF RELY ALARM:**Max support 4 channel rely alarm  
output,contact capacity 3A@220VAC/1A@30VDC,Can be  
configured on the High High Limit/ High Limit/Low Limit/Low  
Low Limit alarm

### ■ COMMUNICATION PRINTER

**COMMUNICATION INTERFACE:**RS232C or RS485, support  
Modbus RTU protocol, baud rate----( 1200、2400、9600、19200、  
38400、57600bps)

**PRINT INTERFACE:**RS232C contact with the micro printer  
directly, baud rate ----1200bps

### ■ RECORDING FUNCTION

**RECORDING CAPACITY:**64/128/192/248MB  
(FLASH capacity can choose)

**RECORDER INTERVAL:**1sec to 240secs, divided into 11  
levels: 1s/2s/4s/8s/12s/24s/36s/60s/120s/180s/240s can be

chosen

**TIME STORAGE LENGTH:** The length of time data storage is related to the Flash capacity. Input point, recorder interval, the computation formula is as follows: (THE NUMERICAL UNITS to be consistent with the formula):

$$\text{TIME STORAGE LENGTH} = \frac{\text{FLASH CAPACITY(MB)} \times 1024 \times 1024 \times \text{INTERVAL(SEC)}}{\text{CHANNELS} \times 16 \times 24 \times 3600} (\text{DAY})$$

#### ■ DATA BACKUP

**U DISK:** Various data can be saved to the U disk. The capacity is from 256MB to 32GB

#### ■ POWER SUPPLY

**POWER:** 220VAC, 50HZ. Support 24VDC (22VDC-28VDC) power supply, support 12VDC (11.2VDC-16VDC) power supply

**NOTE:** DC POWER SUPPORT NEED DECLARE WHEN PLACE ORDER

#### ■ PROTECTION FUNCTION

**POWER OFF PROTECTION:** internal memory protect parameters and historical data forever even if power off.

**CLOCK PROTECTION:** Integrated hardware clock, run accurately when power off

#### ■ ERROR PRECISION

**THERMOCOUPLE CJC ERROR:**  $\pm 2^{\circ}\text{C}$

**CLOCK ERROR:**  $\pm 2$  secs/day

#### ■ WORKING ENVIRONMENT

**AMBIENT TEMPERATURE:**  $0 \sim 50^{\circ}\text{C}$  (avoid direct sun light)

**AMBIENT HUMIDITY:**  $0 \sim 85\% \text{R.H.}$  (non-condensing)


#### ■ NET WEIGHT

**WEIGHT:** Approx 1.0 Kg

**Note:** Technical index is the universal index of the instrument, Functional configuration pls refer to the real machine

## CHAPTER 2 INSTALLATION AND WIRING

### INSTRUMENT DIMENSION



### TERMINAL CONNECTION (REFER TO THE OBJECT)




Figure 2-2

### AC INPUT POWER WIRING




Figure 2-3

**Note:** The supply power of the Engineering field must be in the limit of the withstand voltage of the instrument.

### INPUT/OUTPUT SIGNAL WIRING




Figure 2-4

### TRANSMITTER WIRING




Figure 2-5

## ■ RELAY CONTACT OUTPUT WIRING



Figure 2-6

Note: Connection wire cross-sectional area:  $0.5 \sim 2.5 \text{ mm}^2$ , torque:  $50 \text{ Nm}$ . The instrument relay is normally open output mode.

## ■ COMMUNICATION INTERFACE DEFINITION AND WIRING




Figure 2-7

## CHAPTER 3 USAGE PATTERNS AND OPERATION

### ■ 『KEYBOARD FUNCTION』

Sign	Name	Description
	Key Menu	Used to switch operation screens or decimal places, etc.
	Key Left	Used to switch the channel or move the cursor forward, etc.
	Key Right	Used to switch the channel or move the cursor backward, etc.
	Key Up	Used to switch or increase the value, etc.
	Key Down	Used to switch or decrease the value, etc.
	Key Enter	Used to activate the input box
	Key Combination	Used to switch from operation mode to setting mode (Pressing for 1 second)

Figure 3-1

### ■ 『RUN MODE』




Figure 3-2

## ■ 『OVERVIEW SCREEN』




Figure 3-3

- ◆ Real-time value: display the engineering data of the channel engineering.
- ◆ Unit: display the engineering unit, can free configuration.

## ■ 『DIGITAL DISPLAY』




Figure 3-4

- ◆ Bar chart: filled area shows the present measurement percent of the range.
- ◆ Alarm status: From top to bottom respectively is upper limit/limit/low limit/lower limit, ☐ Normal condition, ☒ Transfinite alarm
- ◆ Relay output status: Relay output display, ② shows k2 relay action is in output alarm status, ① shows K1 relay no action status.
- ◆ Circulation state: in system configuration open auto switch, means switching cycle each channel's digital screen, Switching

interval is adjustable; No auto switch, no display.

## ■ 『CURRENT TREND』




Figure 3-5

- ◆ Time scale: 2M indicate the 2mins curve-segment.
- ◆ Real-time curve: Measure channel/computation data is correspondingly to the left end of the curve.
- ◆ Ruler: The weight scale of the screen curve, is correspondingly to the left screen percentage of the current screen and the component can be adjustable.

## ■ 『HISTROY TREND』




Figure 3-6

- ◆ **Function tag:**The present recall position,when the cursor is on  $\Rightarrow$ ,press **【Enter】** hidden/display cursors;press **【Up】** or **【Down】** realize continuous and rapid recall history curve.
- ◆ **Fixed point recall:**When the cursor is at the position of the recall time,press **【Up】** or **【Down】** adjust the number then press **【Enter】** can realize fixed point recall.

■ [PRINT SCREEN]

Printing configuration:[print] for communication mode, in accordance with the 『RUN MODE』 operating switch:


BEG 11-05-28 18:55:03	Timerprint Yes
END 11-05-28 19:55:03	Interval 1 Min
Chnl 01 Interval 01	Chnl 01-04
Type Data Timerprint	ManualPrt Printing
 <b>Print</b>	
ManualPrint	Timerprint

Figure 3-7

- ◆ Begin/end time: Begin time < end time < clock time.
- ◆ Print/stop: manual start/stop button.
- ◆ Channel/type: manual printing only print a single channel or historical data or curve, Timing printing can print multi-channel current data at the same time. Type be data and curve.
- ◆ Timing/manual printing button: manual switch/timing printing link button.
- ◆ Manual printing interval: the interval is according to unit recording interval.
- ◆ Regular print interval: regular print interval, the minimum interval of 1 min.
- ◆ Timed open print button: [open] when open, auto printing

after opening,no manual.

## ■ 『ALARM INFORMATION』

[illegible]

Figure 3-8

- ◆ Alarm index: Max keep 24 groups alarm information, single channel can show 3 group information at most
- ◆ Alarm/Undo alarm time: Line first is alarm time, second line is Undo alarm time, no Undo alarm time is --/-- --:--:--
- ◆ Alarm types: High High Limit (HH)、High Limit (HI)、Low Limit (LO)、Low Low Limit (LL)
- ◆ Channel number: Current alarm information channel number
- ◆ Scrollbar: Current page

■ 『POWER INFORMATION』

Index	Power-down time	Power-up time
03	11/05/26	09:42:46
	11/05/26	09:45:07
04	11/05/26	17:38:25
	11/05/27	08:51:36
05	11/05/27	16:17:11
	11/05/27	16:29:51

Figure 3-9

- ◆ Power-down index:Keep most 24 power-down/up information
- ◆ Power-down/up time:First line is power-down time, second line is power-up time
- ◆ Scrollbar:Current page

#### ■ 『BACKUP HISTORY DATA』

```

Chnl 01-04      On line
BEG 11-11-03   10:15:03
END 11-11-03   13:17:48
File 1103125.BIN
  ██████████ Backup
  
```

Figure 3-10

- ◆ Equipment:Show mode of U disk,on line,off line,fault state,can not detect U disk[off line],error in the back processing[fault]
- ◆ Channel:The number of backup channel, 01-01 indicate backup 1 channel, 01-04 indicate 1-4 channel need backup
- ◆ Begin: Need backup the start time of historical data
- ◆ End:Need backup the end time of historical data
- ◆ File:backup file name
- ◆ Progress bar: Show the current progress of the backup process
- ◆ Backup/reset/complete:Start, reset, backup complelet status button
- ◆ When the cursor is in the [BEG]or [END]Settings box, press 【Enter】 into quickly jump
- ◆ Such as an error in the way or need to copy again,put the cursor on [backup]button,press 【Up】 or 【Down】 choose to be[reset],press 【Enter】 then can into detecting the storage

device again, when backup finish,[backup] or [reset] button come to be [complete],if need continue copy put the cursor on the [complete] button and press 【Up】 or 【Down】 to choose [backup] and press 【Enter】

## CHAPTER 4 CONFIGURATION AND ASISTING OPERATION

### ■ 『CONFIGURATION』 AND 『SYSTEM CONFIGURATION』

```

Password:00 00 00*
System      Input
Output      Exit
  
```

Configuration screen

```

Time 11-05-28 19:55:03
Password 00 00 00
Interval 240 S Comset
Switch 00 S SYS-MAINT
CJC 28 °C Exit
  
```

System configuration

Figure 4-1

- ◆ Hold on 【Menu】 and 【out】 1 sec later,enter in 『configuration』 screen
- ◆ 『Configuration』 innitial password is [00 00 00],after enter in the correct password press【Enter】into classification menu, '\*' no disappear indicate erro password or no press 【Enter】 in confirm
- ◆ Time setting:System date and time set
- ◆ User password:User password authority set
- ◆ Record interval:Can be 1/2/4/8/12/24/36/60/120/180/240s.The longer the record interval is, the longer of record time, or the short the record interval is, the short of the record time, Usual, the rapid change of the measured signal, need choose the short record interval. Or opposite,when the measured signal changes slowly, can choose the long record interval.

- ◆ Auto switch: Use to auto switch current curve digital screen channel, start with 0S, indicate not auto switch, can adjust the numbers on secondary screen rapidly
- ◆ Cold junction compensation (CJC): has setting and external two mode. When cursor at [CJC], press **【Enter】** into [Cold setting] configuration to choose the two mode, [setting] the code data under setting can be adjustable, [external] the cold data mode can catch by auto, system is under external mode
- ◆ System maintenance: Use to default Settings/ clear alarm information/power information and so on

## ■ [COMMUNICATION CONFIGURATION]

In [system configuration] the [communication configuration] can set the different on line mode:

Mode PC Addr 001	Mode Print Addr 001
Stopbits 2 Parity Odd	
Baudrate 38400	Baudrate 1200
<b>Exit</b>	<b>Exit</b>

Figure 4-2

- ◆ Address: Communication address is a use of identify in the network, it is the identification of the instrument in network; pc software use this to access the instrument; in a same communication address the ip address can be setted in 1-255, can not be repeat
- ◆ Stopbits and parity: need the same setting of the pc software, default setting
- ◆ Baud rate: PC mode baud rate can be changed (must the same to pc software), Printer mode baud rate can't be changed

## ■ [INPUT CONFIGURATION]

Chnl 01	Type Pt100
Alarm	Unit °C
Range -99.9 ~ 850.0	
Adj. 0.0	Filter 0 S
Copy	Paste <b>Exit</b>

Figure 4-3.1

- ◆ Channel  
According to customer demand, 4 is the most
- ◆ Type  
It support several type message, and the analog signal support for universal input, change different signal types without change module, only need change the terminal connection and set the right signal type. Instrument support a variety of signal types such as: 0~20mV, 0~100mV, 0~5V, 1~5V, 0~10mA, 0~20mA, 4~20mA, Pt100, Cu50, Cu100, G53, BA1, BA2, B, E, N, R, J, K, S, T, WRe3-25, WRe5-26, F1, F2, PI, Res (resistance)
- ◆ Unit  
The instrument has the following engineering unit for choose: m3/h, m3/min, m3/s, L/h, L/min, L/s, t/h, t/min, t/s, kg/h, kg/min, kg/s, °C, Pa, kPa, MPa, kgf/cm<sup>2</sup>, Bar, mmH<sub>2</sub>O, mmHg, %, ppm, r/min, ph, mm, Hz, kHz, mA, A, kA, mV, V, kV, VA, kVA, W, kW, MW, J, kJ, kg, kWh and so on
- ◆ Measurement range  
User-defined range scope, the high limit and low limit, can

configure freely. Press **【Menu】** to adjust the decimal digits of measurement range, press **【Enter】** into **〔secondary screen〕** to rapid change the parameters

◆ Alarm setting

HH: 850	Relay: None
HI: 735	Relay: 02
LO: 46	Relay: 01
LL: -99.9	Relay: 02
Zone: 10.0%	<b>Exit</b>

Figure 4-3.2

①**Alarm threshold:**HH、HI、LO、LL is refer to high high limit, high limit, low limit, low low limit alarm, the value under is correspondingly to the alarm shreshold, when actual value bigger than correspondingly limits, differential and difference, can appear or eliminate alarm signal.

②**Alarm contactor:**relay serial,such as contactor 01 is No.1 relay.When value of signal out of setting, corresponding contactor will have an action, close contactor. Choose "No"indicate no matter signal value over or not, the relay keey no action, still in alarm list has alarm data.

③**Alarm differntcial:**alarm differentcial is to avoid instrument repeat alarm at alarm edge, such as range at 0~100, upper alarm dot is80, lower alarm dot is 20, alarm difference is 5.0%, when happen high alarm, the range of instrument is less 80-100\*5%=75, then cancel the alarm this time, the same, when happen lower alarm, the data range at 20+100\*5%=25, then cancell the alarm this time.

◆ Adjust

Allow the user adjust the deviation value, value =

measurement value +adjustable value, normally, set the adjustable value at 0

◆ Filter

Filtering time constant setting, Filter setting method:

$$\text{Show value} = \frac{\text{Last value} \times \text{Input filter} + \text{Current value}}{\text{Input filter} + 1}$$

◆ Copy and paste function

If the setting of each channel is the same, after one channel setting press cursor to [copy], press enter, to copy next setting, press cursor to "channel" setting box, to change the channel, then press cursor to [paste],press enter can copy the function setting of the last into every setting box

■ **〔OUTPUT CONFIGURAION〕**

Chnl 04	Type AO
Action	Reverse
Range	0.0 ~100.0
Source	CH3
	<b>Exit</b>

Figure 4-4

◆ **Action:** Under direct of AO **〔Analog Output〕**,Transmitting range high limit to the output 20mA, Transmitting range low limit of output 4 mA; Under reverse of AO, Transmitting range high limit of output 4 mA,Transmitting range low limit of output 20mA.

◆ **Signal source:**setting the transmitting output signal source channel, CH3 is the third channel.

configure freely. Press[Menu] to adjust decimal digits of measurement range, press[Enter] into [secondary screen] to rapid change the parameters

to rapid change the parameters

#### ◆ Alarm setting

HH: 850	Relay: None
HI: 735	Relay: 02
LO: 46	Relay: 01
LL: -99.9	Relay: 02
Zone: 10.0%	<b>Exit</b>

Figure 4-3.2

①**Alarm threshold:**HH、HI、LO、LL is refer to high high limit, high limit, low limit, low low limit alarm, the value under is correspondingly to the alarm shreshold, when actual value bigger than correspondingly limits, differential and difference , can appear or eliminate alarm signal.

②**Alarm contactor:**relay serial,such as contactor 01 is No.1 relay.When value of signal out of setting, corresponding contactor will have an action, close contactor. Choose "No"indicate no matter signal value over or not, the relay keey no action, still in alarm list has alarm data.

③**Alarm differntcial:**alarm differential is to avoid instrument repeat alarm at alarm edge, such as range at 0~100, upper alarm dot is80, lower alarm dot is 20, alarm difference is 5.0%, when happen high alarm, the range of instrument is less  $80-100*5\%=75$ , then cancel the alarm this time, the same, when happen lower alarm, the data range at  $20+100*5\%=25$ , then cancell the alarm this time.

#### ◆ Adjust

Allow the user adjust the deviation value, value =

measurement value +adjustable value, normally , set the adjustable value at 0

#### ◆ Filter

Filtering time constant setting, Filter setting method:

$$\text{Show value} = \frac{\text{Last value} \times \text{Input filter} + \text{Current value}}{\text{Input filter} + 1}$$

#### ◆ Copy and paste function

If the setting of each channel is the same, after one channel setting press cursor to [copy], press enter, to copy next setting, press cursor to "channel" setting box, to change the channel, then press cursor to [paste],press enter can copy the function setting of the last into every setting box

### ■ [OUTPUT CONFIGURAION]

Chnl 04	Type AO
Action	Reverse
Range	0.0 ~100.0
Source	CH3
	<b>Exit</b>

Figure 4-4

◆ **Action:** Under direct of AO【Analog Output】 ,Transmitting range high limit to the output 20mA, Transmitting range low limit of output 4 mA; Under reverse of AO, Transmitting range high limit of output 4 mA,Transmitting range low limit of output 20mA.

◆ **Signal source:**setting the transmitting output signal source channel, CH3 is the third channel.

## ■ [SECONDARY SCREEN]

5,873	5186
Range:0.000~9.999	Error! Out of range
Enter Cancel	Enter Cancel

Figure 4~5

Secondary screen is mainly use to auto switch setting.Cold, junction compensation setting. Alarm limit setting, measurement, range limit setting and so on. Enter into secondary screen can, quickly adjust multibit vlue, once the value is beyond the scope, the system will note you it is out of parameter range, key operate, please refer to [keyboard function]